

All data taken at the Pacific Northwest National Laboratory
FTS Operators: Tyler O. Danby and John S. Loring
Data Analysis: Russell G. Tonkyn and Matthew S. Taubman

Composite spectrum for: tert-Butyl methyl ether

- First Column: Position in wavenumber (cm^{-1})
- Second column: Real refractive index $n(\tilde{\nu})$ (dispersion index)
- Third column: Imaginary refractive index, $k(\tilde{\nu})$ (absorption index per unit length in centimeters)

Where the complex refractive index $\hat{n} = n(\tilde{\nu}) + ik(\tilde{\nu})$

Following Bertie (in the references below) we define the absorbance as $A = -\log_{10}(I/I_0)$ and the linear absorption coefficient $K = A/d$, where d is the path length. The connection between the imaginary refractive index and the absorbance coefficient arises from the following: $2.303K = 4\pi\tilde{\nu}k$

See the following references for a detailed description of terms and units:

- 1) Bertie, J. E., Zhang, S. L., Eysel, H. H., Baluja, S., & Ahmed, M. K. (1993). Infrared Intensities of Liquids XI: Infrared Refractive Indices from 8000 to 2 cm^{-1} , Absolute Integrated Intensities, and Dipole Moment Derivatives of Methanol at 25°C . *Applied Spectroscopy*, 47(8), 1100-1114. doi:10.1366/0003702934067973
- 2) Bertie, J. E., Zhang, S. L., & Keefe, C. D. (1995). Measurement and use of absolute infrared absorption intensities of neat liquids. *Vibrational Spectroscopy*, 8(2), 215-229. doi:10.1016/0924-2031(94)00038-i

Sample:

- Chemical name, formula and CAS number: tert-Butyl methyl ether, $\text{C}_5\text{H}_{12}\text{O}$, [1634-04-4]
- IUPAC name: 2-Methoxy-2-methylpropane
- Synonyms: Methyl t-butyl ether; Methyl tert-butyl ether; MTBE
- Physical properties: FW = 88.15 g/mole; mp = -108.6°C ; bp = $55-56^\circ\text{C}$; $\rho = 0.74\text{ g/cm}^3$
- Supplier and stated purity: Sigma-Aldrich, anhydrous, 99.8% (Lot # SHBG7709V)
- Temperature of sample: 26°C ($\pm 1^\circ\text{C}$)
- Individual samples were measured at the following path lengths: 3.91, 4.11, 4.73, 6.99, 15.3, 37.3, 99.3, 207, 518 and 1087 micrometers (μm). Final data are a composite of these spectra.
- Sample cell window material is potassium bromide (KBr).
- Preparation: None.

Instrument Parameters:

- Bruker Tensor 27 FTIR, purged with UHP nitrogen
- Spectral range: 7800 to 400 cm^{-1} (1.282 to 25 microns)
- Instrument resolution: 2.0 cm^{-1}
- Number of interferograms averaged per single channel spectrum: 128
- Apodization: Norton-Beer, Medium
- Phase correction: Mertz
- Scanner velocity: 10 kHz
- Folding limits: 15802 to 0 cm^{-1}
- Interferogram zerofill: 4x
- Spectral interval after zerofilling: 0.4823 cm^{-1}
- IR source: Silicon carbide glow bar
- Beamsplitter: Broadband potassium bromide (KBr)
- Detector: DLTGS at room temperature
- Aperture: 3 mm

Measured Refractive Index:

The refractive index for tert-Butyl methyl ether was measured at 15.8 °C and 4.5 °C using an Atago model DR-M2/1550 Abbe refractometer. Notch filters were employed in front of a white light source to make measurements at multiple wavelengths. An InGaAs camera was used to detect signal at 1550 nm. The temperature was controlled to match that in the sample compartment of the FTIR using a heated circulating bath. The measurements at 15.8 °C were as follows:

480 nm: $n = 1.3754$	486 nm: $n = 1.3753$	546 nm: $n = 1.3725$
589 nm: $n = 1.3710$	644 nm: $n = 1.3695$	656 nm: $n = 1.3692$
1550 nm: $n = 1.3623$		

The refractive index, n , vs. wavelength in microns, λ , was fit to an equation similar to that of Sellmeier:

$$n(\lambda) = \{a + b/(\lambda^2 - c)\}^{1/2}$$

The resulting best-fit equation was used to find the refractive index at the highest energy data point in our experimental spectra. For tert-Butyl methyl ether, the result was

$$n(7800 \text{ cm}^{-1}) = 1.3630 \text{ at } 15.8 \text{ }^\circ\text{C}.$$

The change in refractive index with increasing temperature (dn/dT) was estimated to be $-0.00058/^\circ\text{C}$ using the measured values from 4.5 °C and 15.8 °C. Using this estimate of dn/dT gives a projected value for tert-Butyl methyl ether of

$$n(7800 \text{ cm}^{-1}) = 1.3565 \text{ at } 27 \text{ }^\circ\text{C}.$$

a)



b)



Figure 1: The Bruker Tensor 27 FTIR (a) and Abbe refractometer (b).

Post Processing and Related Parameters:

A composite spectrum was created from 10 absorbance spectra (base-10) taken at 10 path lengths: 3.91, 4.11, 4.73, 6.99, 15.3, 37.3, 99.3, 207, 518 and 1087 micrometers (μm). At each path length several spectra were measured and the results averaged for better signal to noise. The measured cell lengths were adjusted using Beer's law plots.

- 1) The imaginary part of the refractive index, or k vector, was determined for each absorbance file as per Bertie's program "RNJ46A" (see reference above). This takes into account the reflective losses due to the KBr windows.

- 2) A composite k vector is created via a classical, weighted, linear, least squares fit using the output files of program "RNJ46A": Intercept=0, slope is fitted, individual absorbance values weighted by T^2 (transmission squared), all absorbance values ≥ 2.5 are given zero weight. Five composite vectors were created and merged by hand.
- The first k vector used the results from the 1087 and 518 μm cells. This k vector determined the final values for the range from 7800 to 3038 cm^{-1} and 2730 to 1540 cm^{-1} .
 - The second k vector used the results from the 207 and 99.3 μm cells. This k vector determined the final values for the range from 3038 to 3001 cm^{-1} , 1540 to 1478 cm^{-1} , 530 to 463 cm^{-1} and 452 to 400 cm^{-1} .
 - The third k vector used the results from the 37.3 and 15.3 μm cells. This k vector determined the final values for the range from 3001 to 2993 cm^{-1} , 1478 to 1094 cm^{-1} , 1081 to 530 cm^{-1} and 463 to 452 cm^{-1} .
 - The fourth k vector used the results from the 6.99 and 4.73 μm cells. This k vector determined the final values for the range from 2993 to 2730 cm^{-1} .
 - The fifth k vector used the results from the 4.11 and 3.91 μm cells. This k vector determined the final values for the range from 1094 to 1081 cm^{-1} .
- 3) The resulting composite k vector and the refractive index at 7800 cm^{-1} were used to create the real or n vector using the Kramers-Kronig relation, as per Bertie's program "LZZKTB."
- Calculated and estimated errors: Type A = 0.7%.
 - Frequency correction (already applied): $\tilde{\nu}(\text{corrected}) = [\tilde{\nu}(\text{instrument}) * .99977 - .01872]$ as determined by comparing measured atmospheric spectral lines (H_2O and CO_2) to values from the Northwest Infrared Spectral Library Database.
 - Axis units: X = Wavenumbers (cm^{-1}); Y = Absorbance (base 10).

Photograph of Sample tert-Butyl methyl ether:



Figure 2: Tert-Butyl methyl ether in Sigma-Aldrich container.